

Anacon, Inc.

Date: 21-Feb-08

CLIENT: Texas Water Development Board
 Lab Order: 0710246
 Project:
 Lab ID: 0710246-001

Client Sample ID: 45-41-701 (3600)
 Collection Date: 10/24/2007 11:15:00 AM

Matrix: WATER

Analyses	Result	Limit	Qual	Units	DF	Date Analyzed
ICP-MS METALS		E200.8				Analyst: HN
Aluminum	2.68	0.700		µg/L	1	11/5/2007 2:54:00 PM
Antimony	ND	0.836		µg/L	1	11/5/2007 2:54:00 PM
Arsenic	ND	0.733		µg/L	1	11/5/2007 2:54:00 PM
Barium	15.4	0.614		µg/L	1	11/5/2007 2:54:00 PM
Beryllium	ND	0.835		µg/L	1	11/5/2007 2:54:00 PM
Boron	298	3.00		µg/L	1	11/5/2007 2:54:00 PM
Cadmium	ND	0.654		µg/L	1	11/5/2007 2:54:00 PM
Chromium	ND	1.17		µg/L	1	11/5/2007 2:54:00 PM
Cobalt	1.60	0.593		µg/L	1	11/5/2007 2:54:00 PM
Copper	ND	0.904		µg/L	1	11/5/2007 2:54:00 PM
Iron	7.33	0.739		µg/L	1	11/5/2007 2:54:00 PM
Lead	ND	0.843		µg/L	1	11/5/2007 2:54:00 PM
Lithium	130	3.00		µg/L	1	11/5/2007 2:54:00 PM
Manganese	19.0	0.137		µg/L	1	11/5/2007 2:54:00 PM
Mercury	ND	1.14		µg/L	1	11/5/2007 2:54:00 PM
Molybdenum	ND	0.956		µg/L	1	11/5/2007 2:54:00 PM
Selenium	1.18	0.989		µg/L	1	11/5/2007 2:54:00 PM
Silver	ND	0.886		µg/L	1	11/5/2007 2:54:00 PM
Strontium	9290	3.00		µg/L	1	11/5/2007 2:54:00 PM
Thallium	ND	0.363		µg/L	1	11/5/2007 2:54:00 PM
Uranium	ND	1.00		µg/L	1	11/5/2007 2:54:00 PM
Vanadium	ND	2.55		µg/L	1	11/5/2007 2:54:00 PM
Zinc	2.02	0.679		µg/L	1	11/5/2007 2:54:00 PM
ICP-MS METALS MODIFIED		E200.8				Analyst: HN
Calcium	152	0.04		mg/L	1	10/31/2007 2:57:00 PM
Magnesium	207	0.13		mg/L	1	10/31/2007 2:57:00 PM
Potassium	15.5	0.17		mg/L	1	10/31/2007 2:57:00 PM
Sodium	500	0.05		mg/L	1	10/31/2007 2:57:00 PM
SILICA AS SiO2		E200.8				Analyst: HN
Silica	46	0.050		mg/L	1	11/5/2007
ANIONS BY ION CHROMATOGRAPHY		E300				Analyst: JL
Bromide	1.61	0.100		mg/L	1	11/9/2007
Chloride	250	0.500		mg/L	1	11/9/2007
Fluoride	2.55	0.500		mg/L	1	11/9/2007
Nitrogen, Nitrate (As N)	ND	0.150		mg/L	1	11/9/2007
Nitrogen, Nitrate / Nitrite (as N)	ND	0.150		mg/L	1	11/9/2007
Sulfate	1900	0.500		mg/L	1	11/9/2007
ANION / CATION BALANCE		A/C BALANCE				Analyst: MK
Anion / Cation Balance	-4.65	5.00		%	1	11/9/2007

Qualifiers:
 ND - Not Detected at the Reporting Limit
 J - Analyte detected below quantitation limits
 B - Analyte detected in the associated Method Blank
 * - Value exceeds Maximum Contaminant Level

S - Spike Recovery outside accepted recovery limits
 R - RPD outside accepted recovery limits
 E - Value above quantitation range

Anacon, Inc.

Date: 21-Feb-08

CLIENT: Texas Water Development Board
Lab Order: 0710246
Project:
Lab ID: 0710246-001

Client Sample ID: 45-41-701 (3600)
Collection Date: 10/24/2007 11:15:00 AM

Matrix: WATER

Analyses	Result	Limit	Qual	Units	DF	Date Analyzed
ALKALINITY			M2320 B			Analyst: SB
Alkalinity, P (as CaCO3)	ND	20.0		mg/L	1	10/31/2007
Alkalinity, Total (as CaCO3)	306	20.0		mg/L	1	10/31/2007

Qualifiers:
ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level

S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range