

Anacon, Inc.

Date: 03-Jul-08

CLIENT: Texas Water Development Board
 Lab Order: 0805007
 Project:
 Lab ID: 0805007-009

Client Sample ID: 24-28-104 (917)
 Collection Date: 4/24/2008 8:30:00 AM

Matrix: WATER

Analyses	Result	Limit	Qual	Units	DF	Date Analyzed
ICP-MS METALS		E200.8				Analyst: HN
Aluminum	ND	0.700		µg/L	1	6/17/2008 9:00:00 PM
Antimony	ND	0.836		µg/L	1	6/17/2008 9:00:00 PM
Arsenic	6.06	0.733		µg/L	1	6/17/2008 9:00:00 PM
Barium	24.1	0.614		µg/L	1	6/17/2008 9:00:00 PM
Beryllium	ND	0.835		µg/L	1	6/17/2008 9:00:00 PM
Boron	280	3.00		µg/L	1	6/17/2008 9:00:00 PM
Cadmium	ND	0.654		µg/L	1	6/17/2008 9:00:00 PM
Chromium	ND	1.17		µg/L	1	6/17/2008 9:00:00 PM
Cobalt	ND	0.593		µg/L	1	6/17/2008 9:00:00 PM
Copper	1.84	0.904		µg/L	1	6/17/2008 9:00:00 PM
Iron	ND	0.739		µg/L	1	6/17/2008 9:00:00 PM
Lead	ND	0.843		µg/L	1	6/17/2008 9:00:00 PM
Lithium	260	3.00		µg/L	1	6/17/2008 9:00:00 PM
Manganese	0.320	0.137		µg/L	1	6/17/2008 9:00:00 PM
Mercury	ND	1.14		µg/L	1	6/17/2008 9:00:00 PM
Molybdenum	2.84	0.956		µg/L	1	6/17/2008 9:00:00 PM
Selenium	35.8	0.989		µg/L	1	6/17/2008 9:00:00 PM
Silver	ND	0.886		µg/L	1	6/17/2008 9:00:00 PM
Strontium	1800	3.00		µg/L	1	6/17/2008 9:00:00 PM
Thallium	ND	0.363		µg/L	1	6/17/2008 9:00:00 PM
Uranium	7.26	1.00		µg/L	1	6/17/2008 9:00:00 PM
Vanadium	31.2	2.55		µg/L	1	6/17/2008 9:00:00 PM
Zinc	4.13	0.679		µg/L	1	6/17/2008 9:00:00 PM
ICP-MS METALS MODIFIED		E200.8				Analyst: HN
Calcium	139	0.04		mg/L	1	6/25/2008 10:54:00 AM
Magnesium	88.2	0.13		mg/L	1	6/25/2008 10:54:00 AM
Potassium	9.49	0.17		mg/L	1	6/25/2008 10:54:00 AM
Sodium	78.0	0.05		mg/L	1	6/25/2008 10:54:00 AM
SILICA AS SiO2		E200.8				Analyst: HN
Silica	78	0.30		mg/L	1	6/25/2008 10:54:00 AM
ANIONS BY ION CHROMATOGRAPHY		E300				Analyst: KK
Bromide	2.56	0.100		mg/L	1	5/20/2008
Chloride	211	0.500		mg/L	1	5/20/2008
Fluoride	3.04	0.500		mg/L	1	5/20/2008
Nitrogen, Nitrate (As N)	2.64	0.150		mg/L	1	5/17/2008
Nitrogen, Nitrate / Nitrite (as N)	2.64	0.150		mg/L	1	5/17/2008
Sulfate	214	0.500		mg/L	1	5/20/2008
ANION / CATION BALANCE		A/C BALANCE				Analyst: MK
Anion / Cation Balance	-4.01	5.00		%	1	6/25/2008

Qualifiers:
 ND - Not Detected at the Reporting Limit
 J - Analyte detected below quantitation limits
 B - Analyte detected in the associated Method Blank
 * - Value exceeds Maximum Contaminant Level

S - Spike Recovery outside accepted recovery limits
 R - RPD outside accepted recovery limits
 E - Value above quantitation range

Anacon, Inc.

Date: 03-Jul-08

CLIENT: Texas Water Development Board
Lab Order: 0805007
Project:
Lab ID: 0805007-009

Client Sample ID: 24-28-104 (917)
Collection Date: 4/24/2008 8:30:00 AM

Matrix: WATER

Analyses	Result	Limit	Qual	Units	DF	Date Analyzed
ALKALINITY			M2320 B			Analyst: SB
Alkalinity, P (as CaCO3)	ND	20.0		mg/L	1	5/5/2008
Alkalinity, Total (as CaCO3)	307	20.0		mg/L	1	5/5/2008

Qualifiers: ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level

S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range